

UK Declaration of Conformity

Manufacturer:	BPR Medical Limited
Address:	22 Hamilton Way Mansfield, Nottinghamshire NG18 5BU, United Kingdom
Name of Device:	Domiciliary Oxygen Switch
Model Numbers:	See appendix 1
Device Classification:	Class IIa (Rule 2)
Applied Common Specifications, harmonised standards and other standards:	BS EN 1041:2008+A1:2013 - Information supplied by the manufacturer of medical devices BS EN ISO 15001:2011 - Anaesthetic and respiratory equipment - Compatibility with oxygen
GMDN Code :	64445
GMDN Term	Oxygen Therapy System

I, the undersigned, hereby declare that the medical device specified above, manufactured on or after the date given below, conforms to the Essential Requirements of Annex I of the UK MDR 2002 (as amended).

UKCA marking is applied the basis of the Annex II route to conformity (full quality assurance). This declaration is supported by UKCA Certificate Number GB22/00000436 issued by SGS United Kingdom Ltd. (Approved Body identification number 0120) and BS EN ISO 13485:2016 Quality Management System Certificate Number GB19/963113 issued by SGS United Kingdom Ltd.

Signed:



Kavnit Kaur
Regulatory Affairs Manager
For and on behalf of BPR Medical Limited

Date of Issue: 27/10/2025

Place of Issue: Mansfield, UK

Appendix 1: Part Numbers, Product Names and UDI-DI Covered by this Declaration**Product Name:** Domiciliary Oxygen Switch

Device Ref.	Name	UDI DI
826-0001	Domiciliary Oxygen Switch (1)	05060274501530
826-0002	Domiciliary Oxygen Switch - Box of 10 Blank Switches	05060274501592
826-0004	Domiciliary Oxygen - Membrane Panel (ON/OFF - Box 20)	05060274501608
826-0005	Domiciliary Oxygen - Membrane Panel (2 Way + OFF - Box 20)	05060274501547
826-0006	Domiciliary Oxygen - Membrane Panel (3 Way + OFF - Box 20)	05060274501615
826-0007	Domiciliary Oxygen - Membrane Panel (4 Way + OFF - Box 20)	05060274501554